

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0056
Location:	LONG BRANCH POLICE DEPT		
Calibration File No.:	00830	Calib. Date:	02/18/2020
Certification File No.:	00831	Cert. Date:	02/18/2020
Linearity File No.:	00832	Lin. Date:	02/18/2020
Solution File No.:	00849	Soln. Date:	07/06/2020
Sequential File No.:	00849	File Date:	07/06/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXN S3-0216
Solution Control Lot:	19060	Expires:	02/11/2021
		Bottle No.:	0684

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:17D		
Control 1 EC	0.094%	21:18D	34.0°C	CONTROL TEST FAILED
Control 1 IR	0.094%	21:18D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	21:19D		
Control 2 EC	0.090%	21:20D	34.0°C	CONTROL TEST FAILED
Control 2 IR	0.092%	21:20D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	21:21D		
Control 3 EC	0.085%	21:22D	34.0°C	CONTROL TEST FAILED
Control 3 IR	0.087%	21:22D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	21:23D		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OLSKI

First Name: GEORGE

MI: T

Signature: 

Badge No.: 336

Date: 07/06/2020