

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0056
Location:	LONG BRANCH POLICE DEPT		
Calibration File No.:	00854	Calib. Date:	07/23/2020
Certification File No.:	00855	Cert. Date:	07/23/2020
Linearity File No.:	00856	Lin. Date:	07/23/2020
Solution File No.:	00859	Soln. Date:	08/11/2020
Sequential File No.:	00859	File Date:	08/11/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXN S3-0217
Solution Control Lot:	19060	Expires:	02/11/2021
		Bottle No.:	0008

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:10D		
Control 1 EC	0.097%	04:11D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	04:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:12D		
Control 2 EC	0.096%	04:13D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	04:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:13D		
Control 3 EC	0.096%	04:14D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	04:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:15D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PAGE

First Name: DANA

MI: L

Signature:



Badge No.: 355

Date: 08/11/2020