

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: LONG BRANCH POLICE DEPT
Serial No.: ARXA-0056
Calibration File No.: 00854
Calib. Date: 07/23/2020
Calib. No.: 00033
Certification File No.: 00855
Cert. Date: 07/23/2020
Cert. No.: 00027
Linearity File No.: 00856
Lin. Date: 07/23/2020
Lin. No.: 00027
Solution File No.: 00862
Soln. Date: 09/04/2020
Soln. No.: 00181
Sequential File No.: 00862
File Date: 09/04/2020

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 19060
Model No.: CU-34
Serial No.: DDXN S3-0217
Expires: 02/11/2021
Bottle No.: 0030

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:24D		
Control 1 EC	0.099%	06:25D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	06:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:26D		
Control 2 EC	0.096%	06:27D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	06:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:28D		
Control 3 EC	0.096%	06:28D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	06:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:29D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GARCIA

First Name: PAUL

MI: J

Signature:  #378

Badge No.: 378

Date: 09/04/2020