

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0056
Location: LONG BRANCH POLICE DEPT
Calibration File No.: 00854 **Calib. Date:** 07/23/2020 **Calib. No.:** 00033
Certification File No.: 00855 **Cert. Date:** 07/23/2020 **Cert. No.:** 00027
Linearity File No.: 00856 **Lin. Date:** 07/23/2020 **Lin. No.:** 00027
Solution File No.: 00868 **Soln. Date:** 10/11/2020 **Soln. No.:** 00184
Sequential File No.: 00868 **File Date:** 10/11/2020

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXN S3-0217
Control Solution %: 0.100% **Expires:** 09/02/2022
Solution Control Lot: 20390 **Bottle No.:** 0115

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:28D		
Control 1 EC	0.099%	16:29D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	16:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:29D		
Control 2 EC	0.098%	16:30D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	16:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:31D		
Control 3 EC	0.097%	16:32D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	16:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:33D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OLSKI

First Name: GEORGE

MI: T

Signature: _____

Badge No.: 336

Date: 10/11/2020