

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0056
Location: LONG BRANCH POLICE DEPT
Calibration File No.: 00802 **Calib. Date:** 09/24/2019 **Calib. No.:** 00031
Certification File No.: 00803 **Cert. Date:** 09/24/2019 **Cert. No.:** 00025
Linearity File No.: 00804 **Lin. Date:** 09/24/2019 **Lin. No.:** 00025
Solution File No.: 00827 **Soln. Date:** 01/25/2020 **Soln. No.:** 00170
Sequential File No.: 00827 **File Date:** 01/25/2020

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXN S3-0217
Control Solution %: 0.100% **Expires:** 02/11/2021
Solution Control Lot: 19060 **Bottle No.:** 0692

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:55S		
Control 1 EC	0.100%	03:56S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	03:56S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:57S		
Control 2 EC	0.099%	03:57S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	03:57S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:58S		
Control 3 EC	0.098%	03:59S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	03:59S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:00S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OLSKI

First Name: GEORGE

MI: T

Signature: 

Badge No.: 336

Date: 01/25/2020