

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0056
Location:	LONG BRANCH POLICE DEPT		
Calibration File No.:	00830	Calib. Date:	02/18/2020
Certification File No.:	00831	Cert. Date:	02/18/2020
Linearity File No.:	00832	Lin. Date:	02/18/2020
Solution File No.:	00845	Soln. Date:	06/22/2020
Sequential File No.:	00845	File Date:	06/22/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXN S3-0216
Solution Control Lot:	19060	Expires:	02/21/2021
		Bottle No.:	0625

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:29D		
Control 1 EC	0.102%	00:30D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	00:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:31D		
Control 2 EC	0.098%	00:32D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:33D		
Control 3 EC	0.099%	00:33D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	00:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:34D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARROLL

First Name: SHANE

MI: M

Signature:

 #363

Badge No.: 363

Date: 06/22/2020