

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0056
Location:	LONG BRANCH POLICE DEPT	
Calibration File No.:	00878	Calib. No.: 00034
Certification File No.:	00879	Cert. No.: 00028
Linearity File No.:	00880	Lin. No.: 00028
Solution File No.:	00899	Soln. No.: 00194
Sequential File No.:	00899	File Date: 06/08/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXN S3-0216
Solution Control Lot:	21090	Expires: 03/17/2023
		Bottle No.: 1398

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:52D		
Control 1 EC	0.099%	05:53D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	05:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:54D		
Control 2 EC	0.099%	05:55D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	05:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:56D		
Control 3 EC	0.097%	05:56D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	05:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARROLL

First Name: SHANE

MI: M

Signature: _____

PO Shane Carroll #363

Badge No.: 363

Date: 06/08/2021