

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0056
Location: LONG BRANCH POLICE DEPT
Calibration File No.: 00904 **Calib. Date:** 07/07/2021 **Calib. No.:** 00035
Certification File No.: 00905 **Cert. Date:** 07/07/2021 **Cert. No.:** 00029
Linearity File No.: 00906 **Lin. Date:** 07/07/2021 **Lin. No.:** 00029
Solution File No.: 00914 **Soln. Date:** 08/06/2021 **Soln. No.:** 00197
Sequential File No.: 00914 **File Date:** 08/06/2021

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXN S3-0217
Control Solution %: 0.100% **Expires:** 03/17/2023
Solution Control Lot: 21090 **Bottle No.:** 1393

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:37D		
Control 1 EC	0.100%	06:38D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	06:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:39D		
Control 2 EC	0.099%	06:40D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	06:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:41D		
Control 3 EC	0.099%	06:41D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	06:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:42D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARROLL

First Name: SHANE

MI: M

Signature:

Shane Carroll #363

Badge No.: 363

Date: 08/06/2021