

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0056
Location: LONG BRANCH POLICE DEPT
Calibration File No.: 00904 Calib. Date: 07/07/2021 Calib. No.: 00035
Certification File No.: 00905 Cert. Date: 07/07/2021 Cert. No.: 00029
Linearity File No.: 00906 Lin. Date: 07/07/2021 Lin. No.: 00029
Solution File No.: 00935 Soln. Date: 11/16/2021 Soln. No.: 00202
Sequential File No.: 00935 File Date: 11/16/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXN S3-0217
Control Solution %: 0.100% Expires: 03/17/2023
Solution Control Lot: 21090 Bottle No.: 1437

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:34S		
Control 1 EC	0.100%	07:35S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:35S		
Control 2 EC	0.100%	07:36S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:37S		
Control 3 EC	0.100%	07:38S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:39S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARROLL

First Name: SHANE

MI: M

Signature:

PO Shane Carroll #363

Badge No.: 363

Date: 11/16/2021