

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0056
Location: LONG BRANCH POLICE DEPT.
Calibration File No.: 00878 **Calib. Date:** 01/12/2021 **Calib. No.:** 00034
Certification File No.: 00879 **Cert. Date:** 01/12/2021 **Cert. No.:** 00028
Linearity File No.: 00880 **Lin. Date:** 01/12/2021 **Lin. No.:** 00028
Solution File No.: 00885 **Soln. Date:** 02/24/2021 **Soln. No.:** 00189
Sequential File No.: 00885 **File Date:** 02/24/2021

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXN S3-0216
Control Solution %: 0.100% **Expires:** 09/02/2022
Solution Control Lot: 20390 **Bottle No.:** 0058

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:35S		
Control 1 EC	0.100%	02:36S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	02:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:37S		
Control 2 EC	0.098%	02:37S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	02:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:38S		
Control 3 EC	0.098%	02:39S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	02:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:40S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SAMOL

First Name: GEORGE

MI: L

Signature:  #330

Badge No.: 330

Date: 02/24/2021