

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0056
Location:	LONG BRANCH POLICE DEPT		
Calibration File No.:	00939	Calib. Date:	12/27/2021
Certification File No.:	00940	Cert. Date:	12/27/2021
Linearity File No.:	00941	Lin. Date:	12/27/2021
Solution File No.:	00964	Soln. Date:	05/29/2022
Sequential File No.:	00964	File Date:	05/29/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXN S3-0216
Solution Control Lot:	21360	Expires:	09/01/2023
		Bottle No.:	0054

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:34D		
Control 1 EC	0.101%	01:34D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	01:34D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:35D		
Control 2 EC	0.099%	01:36D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	01:36D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:37D		
Control 3 EC	0.099%	01:38D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	01:38D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:38D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GARCIA

First Name: PAUL

MI: J

Signature: 

#378

Badge No.: 378

Date: 05/29/2022