

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: LONG BRANCH POLICE DEPT
Calibration File No.: 00939
Certification File No.: 00940
Linearity File No.: 00941
Solution File No.: 00949
Sequential File No.: 00949

Calib. Date: 12/27/2021
Cert. Date: 12/27/2021
Lin. Date: 12/27/2021
Soln. Date: 02/22/2022
File Date: 02/22/2022

Serial No.: ARXA-0056
Calib. No.: 00036
Cert. No.: 00030
Lin. No.: 00030
Soln. No.: 00208

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 21360

Model No.: CU-34
Serial No.: DDXN S3-0216
Expires: 09/01/2023
Bottle No.: 0237

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:20S		
Control 1 EC	0.099%	10:20S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	10:20S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:21S		
Control 2 EC	0.096%	10:22S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:22S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:23S		
Control 3 EC	0.097%	10:24S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:24S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: VECCHIONE

First Name: GARY

MI: M

Signature: _____

Badge No.: 343

Date: 02/22/2022